Investigation of Two Double Throws and Three Single Throws Square Surgical Knots – A Preliminary Experimental & Computational Study on Knot Integrity

Amery Chow¹, Kyrin Liong¹, Shujin Lee²

¹Mechanical Design Engineering/Singapore Institute of Technology (University of Glasgow) 10 Dover Drive, Singapore 138683 Kyrin.liong@singaporetech.edu.sg
² Plastic Surgeon, Plastic Surgery by S J Lee/Mount Elizabeth Hospital #16-13 Mount Elizabeth Medical Centre, 3 Mount Elizabeth, Singapore 228510 plasticsurgerybysjlee@gmail.com

Abstract - In this study, the objective was to compare the knot integrity and efficiency between the two most commonly utilized surgical knots – the two double throws (2DT) and three single throws (3ST) knots via experimental testing and computational modelling. A single suture material - Polyglactin 910 (Vicryl; Ethicon, Inc) was selected for this study and all sutures were of a 4-0-gauge size. 12 knotted suture samples (n=12) were prepared for each of the two surgical knot configurations. A tensile testing machine a load cell of 100N was utilized. The knots were loaded via a crosshead movement rate of 1mm/s throughout the test until knot failure, either by suture rupture at the knot or knot slippage. To test knot strength, assuming slippage does not occur, 3D scaled computational models of the 2DT and 3ST knots were created in SOLIDWORKS, and exported to a finite element analysis (FEA) software - ABAQUS. The sutures were subjected to increasing static forces until the yield stress within the suture was achieved. Preliminary results suggest that the 3ST knot has a higher construction time and is more vulnerable to suture rupture when it fails at a value half that of what is required in a 2DT knot, but is more resistant to knot slippage. The 2DT surgical knot, while more vulnerable to slippage, appears to possess relative acceptable strength, with its yield only initiating at a force double of what is observed in a 3ST knot. Therefore, the 2DT knot may be more appropriate in cases where efficiency and resistance to suture rupture is prioritized due to the lower construction time and its higher load at yield. Conversely, where wound dehiscence via slippage carries a high risk, the 3ST that is more resistant to slippage is encouraged, despite the additional time required to construct.

Keywords: Surgical knot integrity; Mechanical experimental testing; Finite element analysis.

1 Introduction

The ideal surgical knot should be able to withstand knot failure either via slippage or rupture [1] in vivo for an appropriate period of time to allow the wound to properly heal [2]. During healing, the suture and invariably the surgical knot, will be subject to various static and dynamic loads as the subject goes about an everyday routine. In the event a knotted suture fails, incisional hernia or surgical wound dehiscence (SWD) may occur [3], which will lead to subsequent complications. Therefore, determining the knot with the best integrity and efficiency is critical.

Although previous studies have investigated the effect of suture techniques and materials experimentally, there remains a paucity of consensus in the outcomes. A study by Riboh et al. [2] concluded that surgeon's knots are stronger than sliding knots. Conversely, Gandini et al. [4] reported that efficiency of the sliding knots over its counterpart. Van Rijssel et al. [5] has argued that even with an additional cross at the base, the surgeon's knot held no additional strength benefits over the square knot. With respect to suture materials, a similar lack of consensus exists between studies [5, 6]. Despite the numerous studies conducted, none have sought to compare two of the most commonly utilized knots – the two double throws square knot (2DT), commonly known as the surgeon's knot, and the three single throws square knot (3ST).

Conventionally, the 3ST knot is the preferred knot of choice as it is ostensibly simpler and more commonly accepted, though this is not based on complete scientific data [8]. 2DT surgical knots, on the other hand, seem to offer greater knot security, although it may require a slightly longer construction time. Surgeons typically choose between these two knots based on institutional practice [9] and yet no prior studies have compared the integrity between the two.

To that end, this study aims to investigate and compare the knot integrity between the 2DT and 3ST surgical knots. Specifically, the analysis will focus on each knot's vulnerability to slippage via experimental testing, suture rupture at the knot via computational modelling, and efficiency in terms of construction time, which is pertinent particularly when multiple sutures are necessary.

2 Methodology

2.1 Sample Preparation

A single suture material - Polyglactin 910 (Vicryl; Ethicon, Inc) was selected for this study and all sutures were of a 4-0-gauge size, absorbable synthetic and of multifilament structure. 12 knotted suture samples (n=12) were prepared for each of the two surgical knot configurations. To avoid operator influence, all surgical knots were tied by a single surgeon with proper instrumentation. The 2DT surgical knot, otherwise known as the surgeon's knot, was prepared with a double-wrapped first throw followed by an additional double throw square knot [10] (Fig. 1a). The 3ST surgical knot was prepared with two square knots, in which knot base is crossed only once [11] (Fig. 1b).



Fig.1 Configuration of a) two double throws squared, and b) three single throws squared surgical knots.

Prior to testing, all samples were knotted around a 44.5 mm diameter cylindrical plastic bottle (Fig 2a). The construction time taken to tie each surgical knot was recorded and compared between knot configurations. The time was measured starting when the suture was looped on the underside of the cylinder and the surgeon began the first loop and was stopped after the surgeon cut the last knot's thread.

The tail ends of the knot were cut to a length of 3mm length [12] and knot loops were cut into equal lengths (Fig 2b). Standard force was applied to the tail ends of each sample to test for knot slippage as defined by a binary value where a displacement exceeding 1.8 mm would constitute an unstable knot, and vice-versa [9]. The displacement was measured by the change of length from the tail ends to the knot.



Fig. 2: a) A surgeon's knot tied around a 45.5mm diameter plastic cylinder. b) A final knot sample with 3mm tail ends and a knot loop cut into equal lengths.

2.2 Experimental Setup

All knots were tested within 12 hours of being tied. Knots that untied before testing were not included in the calculation of the mean maximum tensile load at failure. A tensile testing machine (Criterion 42 Universal testing Machine, MTS, Eden Prairie, MN) with a load cell of 100N (Fig 3) was utilized. The loops ends were clamped between the crosshead grips with a non-harsh adhesive applied on their inner surface. The grips were mechanically tightened with the use of a hex key to

minimize slip during testing. To prevent possible damage to the suture samples, the investigator was careful to not exert excessive force whilst tightening. The distance between grips was adjusted to 90 mm - when the suture was just taut.

The knots were then loaded via a crosshead movement rate of 1mm/s throughout the test until knot failure, either by suture rupture at the knot or knot slippage – wherein a knot completely unravels under tensile loads. Force and displacement values were continuously recorded at a sampling data rate of 10Hz. The mode of failure and maximum tensile load at failure were recorded, and the values compared between the two different suture knot configurations.

2.3 Statistical Analysis

Statistical analyses were performed in Minitab 19 Statistical Software (2020, State College, PA: Minitab, Inc) which ensured that the data points collected for failure load via slippage and knot construction time confirmed to a normal distribution via the Kolmogorov and Smirnov test. Unpaired, parametric student's t-tests were performed on the data sets to identify any significant differences between the two knot configurations. A *P*-value < 0.05 was considered statistically significant.



Fig. 3: Criterion 42 Universal tensile testing machine (MTS, Eden Prairie, MN) utilized in the mechanical testing phase

2.4 Finite Element Analysis

To test knot strength with no knot slippage, under the expert guidance of an experienced plastic surgeon, 3D computational models of the 2DT and 3ST surgical knots were drawn within a computer-aided design (CAD) software Solidworks (2019, Dassualt Systèmes Technologies, Providence, RI). Due to the morphological complexity of the knots and to allow for efficient computational convergence, as a preliminary measure, the suture diameter was scaled up to 2.5 mm from the original 0.15 mm diameter of the Polyglactin 910 (Vicryl; Ethicon, Inc), 4-0-gauge sutures. All other measurements in the models were scaled accordingly. Hence, the analysis of the finite element results will focus on relativity of the values and not the absolute values.

As the computational models are idealized, no external factors concerning angle of pull by the surgeon or force differences corresponding to suture lengths was involved. Therefore, only the knot itself needed to be modelled (Fig. 3), resulting in a computationally efficient analysis due to a decreased number of elements within the model. Additionally, the volume of the each knot was recorded.

The 3D models of the knots were then exported to a finite element analysis (FEA) software, ABAQUS (v6.17, Dassualt Systèmes Technologies, Providence, RI) for simulation configuration. As a first approximation, the suture material was assigned with homogenous, isotropic elastic material properties obtained from CES Edupack 2019 (Granta Design, Cambridge, UK), where the Young's modulus was assigned a value of 2040 MPa, and a Poisson's ratio of 0.405. These values were of a similar order to suture properties utilized in previous studies [12, 13].



Fig. 3: 3D CAD models of the a) Two double throws surgical knot, and b) three single throws surgical knot. The bottom of each suture is is fixed, and the top end of the suture is subjected to an applied load.

In an effort to determine the effect of external tensile loads on the knot strength, the bottom end of the suture was assigned a fixed boundary condition, where all degrees of freedom were constrained. The top ends of the suture were then subjected to a static force uniformly distributed over the area. Multiple simulations were conducted to ascertain the tensile force necessary for each knot configurations to reach a yielding value of 55 MPa, as defined by the suture material defined in CES Edupack 2019 (Granta Design, Cambridge, UK). At this yield stress, the suture would be at its elastic limit, thus suggesting impending suture rupture at the knot.

Both surgical knot models were assigned with explicit 4-node linear tetrahedron elements (C3D4). An element seed size of 0.2 was chosen so that the intricate profile of the knots could be accurately meshed without compromising its external morphology. The 2DT and 3ST knot computational models consisted of 580, 612 and 533, 861 elements, respectively.

3 Results

In the current study, knot security was evaluated in terms of (i) knot instability prior to mechanical testing, (ii) knot slippage/rupture during testing, and (iii) idealized suture rupture at the knot via finite element modelling.

24 suture samples were evaluated via mechanical tensile experiments. Prior to mechanical testing, two of the 3ST knot exhibited instability, where their displacement values exceed 1.8mm, and were therefore excluded from testing. One 2DT sample was further excluded due to a mechanical testing error during that specific trial. Of the 21 remaining samples, only one 2DT knot sample failed by knot breakage at a tensile load of 13.941 N, while all remaining samples failed by knot slippage. Ultimately, 10 data points were present for each knot configuration that failed via slippage. Due to the lack of data

concerning knot rupture, knot failure by slippage will be the primary area of comparison between the two knot configurations. The results of the experimental testing and knot construction times are summarized in Table 1.

Table 1: Results of experimental testing			
Surgical Knot	Two double throws square	Three single throws square	<i>P</i> -value
n	10	10	-
Construction time [Standard Deviation (SD)] (s)	19.30 [2.84]	24.89 [5.66]	0.0153
Load to failure via slippage [SD] (N)	1.85 [0.93]	5.57 [1.17]	5.40×10 ⁻⁵

There were significant differences between the construction times and tensile load to failure via knot slippage between the two knot configurations, where in both cases the *P*-value fell below 0.05. While the 3ST knot required a slightly longer construction time, it appeared to slip at a higher tensile load, suggesting that it is more resistant to knot slippage.

Expectedly, the failure load for suture rupture at the knot of 13.941 N far exceeded the tensile load observed for any failure via knot slippage [12]. However, as there was only a single 2DT knot sample that failed by suture rupture, a comparison between the knot figurations based upon failure by suture rupture could not be investigated. Instead, the computational analyses results, which were a reflection of pure tensile load application and would not mechanically result in any slippage, were then utilized to gain insight into each knot's vulnerability to suture rupture.

The FEA results revealed that an applied force of 35.0 N and 16.8 N would elicit a yield stress in a 2DT and a 3ST surgical knot, respectively. The resultant von Mises Stresses in each suture configuration are shown in figure 4. This suggests that assuming no knot slippage, a 3ST knot would first fail by suture rupture at its knot.



Fig. 4: Resultant von Mises Stresses in the two double throws (2DT) square surgical knot and the three single throws (3ST) surgical knot. Both the front and back views are shown of each configuration.

For an efficient comparison, the relevant values for 3ST were expressed as a fraction of the values observed for the 2DT surgical knot. The results are illustrated in figure 5.



Fig. 5: Graph of relative values, where each value is a fraction of the corresponding associated 2DT value, in the categories of load to failure by slippage and suture rupture at knot, and construction time.

4 Discussion

Knot security is one of the most important elements of a tied suture [9]. As a knot is the most vulnerable part of a suture [1,14], it is paramount to ensure that whichever knot is chosen by a surgeon, it remains secure enough to allow the wound to heal sufficiently. Therefore, the objective of the current study is to evaluate two of the most commonly utilized surgical knots – the 2DT or surgeon's knot, and the 3ST surgical knot on the basis of knot security and construction time.

Through experimental testing and computational analysis, we were able to offer a preliminary comparison between the knot integrity of the two knot configurations. The lower the load at failure, the more vulnerable the knot is to succumb to this mode of failure. The 3ST knot appeared to more resistant to knot slippage with a higher failure load, but more vulnerable to suture rupture when it fails at a value half that of what is required in a 2DT knot (Fig. 5), albeit in the idealized case in which a tensile force that is perfectly in line with the longitudinal direction of the suture is applied to its end and slippage does not occur.

The extra throw in the 3ST configuration increases the area of interfacing surfaces, thereby increasing friction within the loop, and possibly leading to its greater resistance to knot slippage. This suggests that the 3ST knot is strong enough to withstand more loads. Therefore, although the current knot has a higher construction time than a 2DT knot (Fig. 5), in instances when surgical wound dehiscence (SWD) via slippage [16] may be life threatening such as in vessel anastomoses, the more time-consuming 3ST surgical knot may be preferred.

The 2DT surgical knot, on the other hand, while more vulnerable to slippage, appears to possess relative acceptable strength, with its yield only initiating at a force double of what is observed in a 3ST knot. This suggests that the 2DT knot may be able to withstand the normal stresses of wounds. Hence, when a large wound is to be closed that requires multiple sutures [17], the more efficient 2DT that can withstand suture rupture more adequately may be a suitable option.

Although this study has provided a basis for discussion on the knot integrity of these two commonly performed surgical knots, the result that each knot configuration is vulnerable to a different mode of failure suggests that indeed more research and improvements are necessary to properly determine the biomechanical strength and characteristics of each knot. A larger sample size of sutures, where available, may have led to more conclusive observations, specifically with respect to suture rupture. The inclusion of skin tissue samples would have also enabled further study on how each knot configuration affects the skin. Lastly, 1:1 CAD modelling and a parametric study on the material properties of the suture assigned in the FEA could further reveal accurate insights relating to the biomechanics of the two knot configurations that could later be used to help guide surgical decisions.

5 Conclusions

In this study, the primary objective was to compare the knot integrity and efficiency between the two most commonly utilized surgical knots – the 2DT and 3ST knots via experimental testing and computational modelling. Although preliminary, the results suggest that that the 2DT knot may be more appropriate in cases where efficiency and resistance to suture rupture is prioritized due to the lower construction time and its higher load at yield. Conversely, where SWD via slippage carries a high risk, the 3ST that is more resistant to slippage is encouraged, despite the additional time required to construct.

Acknowledgements

The authors wish to thank Johnson & Johnson Pte Ltd for their donation of suture material for the purposes of this study, and Mr Emran Poh from the Singapore Institute of Technology for his help with the illustrations for this paper.

References

- [1] R. F. Edlich and W. B. Long III, "Surgical Knot Tying Manual, Third edition," *Covidien*, 2008. www.covidien.com.
- [2] J. C. Riboh, D. S. Heckman, R. R. Glisson, and C. T. Moorman, "Shortcuts in arthroscopic knot tying: Do they affect knot and loop security?," Am. J. Sports Med., vol. 40, no. 7, pp. 1572–1577, 2012, doi: 10.1177/0363546512446676.
- [3] D. C. Bartlett and A. N. Kingsnorth, "Abdominal wound dehiscence and incisional hernia," *Surgery*, vol. 27, no. 6, pp. 243–250, 2009, doi: 10.1016/j.mpsur.2009.05.001.
- [4] M. Gandini, G. Giusto, F. Comino, and E. Pagliara, "Parallel alternating sliding knots are effective for ligation of mesenteric arteries during resection and anastomosis of the equine jejunum," *BMC Vet. Res.*, vol. 10, no. Suppl 1, pp. 1–6, 2014, doi: 10.1186/1746-6148-10-S1-S10.
- [5] E. J. C. van Rijssel, J. Baptist Trimbos, and M. H. Booster, "Mechanical performance of square knots and sliding knots in surgery: A comparative study," *Am. J. Obstet. Gynecol.*, vol. 162, no. 1, pp. 93–97, 1990, doi: https://doi.org/10.1016/0002-9378(90)90828-U.
- [6] D. M. Marturello, M. S. McFadden, R. A. Bennett, G. R. Ragetly, and G. Horn, "Knot security and tensile strength of suture materials.," *Vet. Surg.*, vol. 43, no. 1, pp. 73–79, Jan. 2014, doi: 10.1111/j.1532-950X.2013.12076.x.
- [7] T. M. Muffly, N. Kow, I. Iqbal, and M. D. Barber, "Minimum number of throws needed for knot security," *J. Surg. Educ.*, vol. 68, no. 2, pp. 130–133, 2011, doi: 10.1016/j.jsurg.2010.11.001.
- [8] J. E. Tidwell, V. L. Kish, J. B. Samora, and J. Prud'homme, "Knot security: how many throws does it really take?," *Orthopedics*, vol. 35, no. 4, pp. e532-7, Apr. 2012, doi: 10.3928/01477447-20120327-16.
- [9] E. Silver, R. Wu, J. Grady, and L. Song, "Knot Security- How is it Affected by Suture Technique, Material, Size, and Number of Throws?," J. Oral Maxillofac. Surg., vol. 74, no. 7, pp. 1304–1312, Jul. 2016, doi: 10.1016/j.joms.2016.02.004.
- [10] T. M. Muffly, J. Boyce, S. L. Kieweg, and A. J. Bonham, "Tensile strength of a surgeon's or a square knot," *J. Surg. Educ.*, vol. 67, no. 4, pp. 222–226, 2010, doi: 10.1016/j.jsurg.2010.06.007.
- [11] R. P. Brown, "Knotting technique and suture materials," *BJS (British J. Surgery)*, vol. 79, no. 5, pp. 399–400, May 1992, doi: https://doi.org/10.1002/bjs.1800790507.
- [12] J. G. Thacker, G. Rodeheaver, L. Kurtz, M. T. Edgerton, and R. F. Edlich, "Mechanical performance of sutures in surgery," Am. J. Surg., vol. 133, no. 6, pp. 713–715, 1977, doi: https://doi.org/10.1016/0002-9610(77)90161-1.
- [13] N. P. Ingle, M. W. King, and M. A. Zikry, "Finite element analysis of barbed sutures in skin and tendon tissues," *J. Biomech.*, vol. 43, no. 5, pp. 879–886, 2010, doi: http://dx.doi.org/10.1016/j.jbiomech.2009.11.012.
- [14] A. M. Zihlif, R. A. Duckett, and I. M. Ward, "The Poisson's ratio of ultra-drawn polyethylene and polypropylene fibres using Michelson interferometry," J. Mater. Sci., vol. 13, no. 8, pp. 1837–1840, 1978, doi: 10.1007/BF00548754.
- [15] H. Tera and C. Aberg, "Strength of knots in surgery in relation to type of knot, type of suture material and dimension of suture thread.," *Acta Chir. Scand.*, vol. 143, no. 2, pp. 75–83, 1977.
- [16] K. Ousey, R. Djohan, C. Dowsett, F. Ferreira, T. Hurd, and M. Romanelli, "Surgical wound dehiscence: improving prevention and outcomes," *World Union Wound Heal. Soc. Consens. Doc.*, p. 4, 2018, [Online]. Available: https://pure.hud.ac.uk/en/publications/surgical-wound-dehiscence-improving-prevention-and-outcomes.
- [17] E. L. Howes and S. C. Harvey, "The Strength of the Healing Wound in Relation to the Holding Strength of the Catgut Suture," *N. Engl. J. Med.*, vol. 200, no. 25, pp. 1285–1291, Jun. 1929, doi: 10.1056/NEJM192906202002504.